

Republic of the Philippines  
**VISAYAS STATE UNIVERSITY**  
Visca, Baybay City, Leyte  
SUPPLY & PROPERTY MANAGEMENT

## INSPECTION & ACCEPTANCE REPORT

Supplier: **LABTRADERS, INC.**

IAR No: 25-10-1370

P.O No.: **BAC2-GOODS-25-20-090-A (STF-CO)**

Date: 2025-10-14

Office/Dept: **DOPAC**

**212,000.00**

STF-2024-10-02469

No.	UNIT	DESCRIPTION	Quantity		Date Received	Balance	Remarks
			Per P.O	Actual			
1	pcs	Shower and eyewash	4	4	2025-10-13 #: DR D-1825	0	Completed

**Specification:**

- Stainless Steel Shower & Eyewash Station.
- Unit comes complete with ABS bowl (250 mm diameter) and ABS shower head (250 mm diameter), chemical resistant PP eye wash nozzle aerated to produce a soft, bubbly, gentle spray with auto-open dust cap, built-in stainless filter & stay-open ball valves, shower head comes with stay-open ball valve and pull rod (pull down to open, push up to close). Unit conforms to ISO 9001:2000 standard. Rate of flow is 20 gallons per minute (gpm) at 30 psi (0.3 MPa) at the shower and is reduced to a flow rate of 1.5 to 2.2 liters per minute (lpm) at the eyewash. The applicable water pressure of the unit is at a minimum of 29.0 psi and a maximum of 87.0 psi (0.2 MPa min-0.6 MPa max).

**INSPECTION**

Date Inspected: 10/14/25

- ☒ Inspected, verified and found in order as to quantity and specifications

☐ Others: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**ACCEPTANCE**

Date Received: 10/14/25

☒ Complete

☐ Partial

**NOREVE JEAN M. AGAD**  
Inspector

**DOREEN B. ALBA**  
Head, Supply & Property Management