#### Republic of the Philippines

## **VISAYAS STATE UNIVERSITY**

Visca, Baybay City, Leyte

#### SUPPLY & PROPERTY MANAGEMENT

# **INSPECTION & ACCEPTANCE REPORT**

Supplier:

KRYPTON INTERNATIONAL RESOURCES

SALES & SERVICES INC.

IAR No:

25-11-1594

P.O No.:

BAC2-GOODS-25-24-004 (STF-CO)

Date:

2025-11-13

Office/Dept: CVM

800,000.00

STF-2025-01-00051

No.	UNIT	DESCRIPTION	Quantity		Date	Balance	Remarks
			Per P.O	Actual	Received	Balance	Kemarks
1	unit	Laboratory Fumehood Complete Unit	1	1	2025-11-12 #: DR 30940, 30941/ SI 25529, 25530	0	Completed

#### Specification:

#### Features:

- Microprocessor control system
- · Digital LED display
- · LED lamp for illumination
- · UV lamp for decontamination of the working environment
- · 9-level adjustable airspeed exhaust blower
- · Air velocity memory function in case of power failure
- · Built-in centrifugal blower
- · 5 mm motorized front glass window
- · Active carbon filter to remove toxic fumes
- Alarm function for filter reaching working time of 3500 hours
- · Chemically resistant phenolic resin work table
- Interior material: Highly authentic melamine board, resistant to weak acid and base
- · Exterior material: Cold-rolled steel with anti-bacterial coating
- · Backside air compensation to avoid turbulence in the working area

### Specifications:

- Exterior Size (W × D × H): 1040 × 800 × 2200 mm
- Interior Size (W × D × H): 820 × 670 × 730 mm
- Maximum Opening: 520 mm
- LED Lamp: 8 W × 1
- UV Lamp: Emission of 253.7 nm for efficient decontamination

• Air Velocity: 0.3 m/s to 0.8 m/s

• Work Surface Height: 850 mm

• Exhaust Duct: PVC 4 meters, ? 300 mm

• Noise Level: ? 68 dB

Power Supply:

o AC 220 V ± 10%, 50/60 Hz

o AC 110 V ± 10%, 60 Hz

• Power Consumption: 400 W

• Package Size (W × D × H): 1190 × 990 × 1690 mm

• Gross Weight: 286 kg

#### Inclusions:

· UV Lamp: 2 pieces

· LED Lamp: 1 piece

· Water Tap: 1 piece

· Gas Tap: 1 piece

· Water Sink: 1 piece

• Base Cabinet: 1 piece

• Total Load of 2 Waterproof Sockets: 500 W - 2 pieces

PVC Exhaust Duct (4 meters): 1 piece

· Pipe Strap: 1 piece

· Active Carbon Filter: 1 piece

· Ducting: 15 meters

Warranty: At least 5 years

With free installation and calibration

INSPECTION	ACCEPTANCE
Date Inspected:	Date Recieved:
Inspected, verified and found in order as to quantity and specifications	Complete
Others:	—— O Partial
- XVE)	
NOREVE JEAN M. AGAD	DOREEN B. ALBA
Unspector	Head, Supply & Property Management