

Republic of the Philippines
VISAYAS STATE UNIVERSITY
Visca, Baybay City, Leyte
SUPPLY & PROPERTY MANAGEMENT

INSPECTION & ACCEPTANCE REPORT

Supplier: **LABTRADERS, INC.**

IAR No: 25-07-0803

P.O No.: **BAC2-GOODS-25-14-013 (GF-CO)**

Date: 2025-07-03

Office/Dept: **DA**

91,250.00

GF-2024-11-03090

No.	UNIT	DESCRIPTION	Quantity		Date Received	Balance	Remarks
			Per P.O	Actual			
1	unit	Analytical Balance, BA-N2204	1	1	2025-07-03 #: DR-1760	0	Completed

Specification:

Capacity: 220 g

Readability: 0.1 mg

Repeatability (Standard Deviation): 0.1 mg

Linearity: ± 0.2 mg

Pan Size: 90 mm \varnothing (3.5 inches \varnothing)

Response Time: 3 seconds

Calibration:

Automatic Internal Calibration

Perfect Self Auto Calibration (temperature, time, power on)

Display: Bright LCD with LED backlight

Units of Measure: Gram, Milligram

Tare Range: Full

Operating Temperature: 10°C to 40°C

Dimensions (W x D x H): Approx. 342.5 x 212 x 341 mm

Power Supply: AC Adapter (100V-240V \pm 20%, 50-60 Hz)

Net Weight with Packing: Approx. 8 kg

Warranty: 1 yr

2	pc	Chlorophyll Meter, BCHL-B	1	1	2025-07-03 #: DR-1760	0	Completed
---	----	---------------------------	---	---	--------------------------	---	-----------

Specification:

Measuring Area: Approx. 2mm x 2mm

Measuring Mode: 2-wavelength concentration difference of optical methods

Sensors: Silicon semiconductor photodiode

Display Mode:

Measuring value: 3-digit liquid crystal display

Measuring times: 2-digit liquid crystal display

Min Measuring Interval: <3 seconds

Measuring Range:

Chlorophyll: 0.0-99.9 SPAD

Temperature: -10 to 99.9°C

Accuracy:

Chlorophyll: ± 3.0 SPAD

Temperature: $\pm 0.5^{\circ}\text{C}$

Repeatability:

Chlorophyll: ± 0.3 SPAD

Temperature: $\pm 0.2^{\circ}\text{C}$

Power Supply: 4.2V - 2000mAh rechargeable Li battery

Memory Capacity: 30KB, automatically calculates and shows the average

Packing Size (W x D x H): Approx. 270 x 210 x 140 mm

Gross Weight: Approx. 1.0 kg

Warranty: 1 yr

INSPECTION

ACCEPTANCE

Date Inspected:

Date Recieved:


☒ Inspected, verified and found in order
as to quantity and specifications

☒ Complete

☒ Others: In box

☐ Partial


JO JANE D. ATOK
Inspector

 7/4/25
VIVIAN V. BALBARINO
Head, Supply & Property Management