

Republic of the Philippines
VISAYAS STATE UNIVERSITY
Visca, Baybay City, Leyte
SUPPLY & PROPERTY MANAGEMENT

INSPECTION & ACCEPTANCE REPORT

Supplier: **NOVEAULAB ASIA CORP.**

IAR No: 25-09-1196

P.O No.: **2FB-23-30-18(TF)**

Date: 2025-09-18

Office/Dept: **CVM**

372,000.00

TF-2022-12-02507

No.	UNIT	DESCRIPTION	Quantity		Date Received	Balance	Remarks
			Per P.O	Actual			
1	set	Portable Freezing Unit for 90 Straws-Complete Set, Minitube 15043/0736 Freezing Unit	1	1	2025-07-07 #: DR-009460	0	Completed
Specification:							
2	set	Portable ultrasound scanner for transabdominal pregnancy detection in sows and small ruminants, MINITUBE 23500/1000 ULTRASOUND SCANNER	1	1	2025-09-17 #: DR-009927/SI-7374	0	Completed

Specification:

Product Features:

- **Monitor Size:** 5.0" TFT-LCD
- **Display Modes:** B, B+B
- **Image Gray Scale:** 256 levels
- **Measure Functions:** Distance, circumference, area, volume
- **Detection Depth:** Up to 240 mm
- **Battery Capacity:** 3000 mAh, 7.4V
- **Power Consumption:** 8W
- **Weight (incl. probe):** 854 g
- **Overall Size:** 240 x 117 x 40 mm (L x W x H)
- **Electrical Connection:** 100-240 V

Scope of Delivery:

- 1 Main Device with 3.5 MHz Electronic Sector Convex Probe and 5" TFT Monitor
- 1 Transport Case
- 2 Carrying Straps
- 5 Knurled Screws for Fixing the Carrying Strap to the Main Unit
- 4 Li-ion Rechargeable Batteries (3.7 V/3000 mAh with Protective Circuit, 2 x 2 pcs.)
- 1 Battery Charger
- 1 Mini 4-pin USB Cable to Transmit Pictures
- 2 Repro Jelly 100 ml
- 1 Manual

INSPECTION

Date Inspected:

9/18/14

☒ Inspected, verified and found in order
as to quantity and specifications

☐ Others: _____

ACCEPTANCE

Date Recieved:

9/18/14

☒ Complete

☐ Partial

NOREVE JEAN M. AGAD

Inspector

DOREEN B. ALBA

Head, Supply & Property Management