

Republic of the Philippines
VISAYAS STATE UNIVERSITY
Visca, Baybay City, Leyte
SUPPLY & PROPERTY MANAGEMENT

INSPECTION & ACCEPTANCE REPORT

Supplier: **QUETAN MEDICAL SUPPLIES TRADING** IAR No: 25-10-1295
P.O No.: **PO-TF-2025-09-0647** Date: 2025-10-06
Office/Dept: **UHS** **24,970.00** TF-2025-07-00915

No.	UNIT	DESCRIPTION	Quantity		Date Received	Balance	Remarks
			Per P.O	Actual			
1	set	Zoll CPR D-pads	1	1	2025-10-03 #: DR 000138/ CI 01002	0	Completed

Specification:

1. ZOLL CPR-D Padz Product : Model Number: ZOLL CPR-D Padz

Shelf Life: 5 years

Conductive Gel: Polymer Hydrogel

Conductive Element: Tin

Packaging: Multi-layer foil laminate pouch

Impedance Class: Low

Cable Length: 36 in (91 cm)

Complete assembly:

Folded length: 7.6 in (19.4cm)

Folded width: 7.0 in (17.8cm)

Folded height: 1.5 in (3.8cm)

Design standards: Meets application requirements of ANSI/AAMI/ISO DF-39-1993

*Unique one-piece design

*Easy to set up: provides all accessories required for rescue session.

*Maximizes proper placement with fast "release and press" liners.

*Virtually impossible for rescuer to get confused

*Not available on any other AED-world's first AED to offer this feature.

*Integrates important CPR skills to help rescuer through all aspects of a rescue.

2. Duracell CR123 Lithium Batteries (Pack of 10)

*Pack of 10 - Specifically for ZOLL AED Plus units

*Lithium IEC: CR17345

*Volt: 3.0

*Size (mm): 34 x 17

*mAh: 1600

*anticipated Life - 5 years in normal standby mode.

3. ZOLL M2 (Monitor/Defibrillator) 5-Lead ECG Cable

*5-Lead Electrode 12Pin

AHA Color Coding

RA/White Electrode

LA/Black Electrode

LL/Red Electrode

RL/Green Electrode

V/Brown Electrode

INSPECTION

ACCEPTANCE

Date Inspected: 10/6/25

Date Recieved: 10/6/25

☒ Inspected, verified and found in order
as to quantity and specifications

☒ Complete

☐ Others: _____

☐ Partial


NOREVE JEAN M. AGAD
Inspector


DOREEN B. ALBA
Head, Supply & Property Management