

Republic of the Philippines
VISAYAS STATE UNIVERSITY
Visca, Baybay City, Leyte
SUPPLY & PROPERTY MANAGEMENT

INSPECTION & ACCEPTANCE REPORT

Supplier: **NOVEAULAB ASIA CORP.**
P.O No.: **VSU-GOODS-BAC2-24-64-008 (TF)**
Office/Dept: **NARC**

IAR No: 25-07-0837
Date: 2025-07-08
177,975.00 TF-2024-07-01415

No.	UNIT	DESCRIPTION	Quantity		Date Received	Balance	Remarks
			Per P.O	Actual			
1	unit	Water Distillation unit, OFFER: LASANY LPH-4/UNIT	3	3	2025-07-07 #: DR-009190/SI-6837	0	Completed

Specification:

- **Output Capacity:** Minimum of 4 liters per hour
- **Power Input:** 3 kW
- **Heater:** Chromium-plated heater
- **Boiler:** Horizontal borosilicate glass boiler
- **Condenser:** High-efficiency
- **Distillate Temperature Range:** 25°C to 40°C (under room temperature)
- **Mounting:** Includes materials for wall mounting
- **Water Quality:** Capable of producing pyrogen-free water
- **Application:** Ideal for general laboratory use

This unit is designed for efficient distillation, ensuring the production of high-quality distilled water for laboratory applications.

Dimensions: 50cmx 15cmx 45cm(WxDxH) to be able fit in the space allocated for installation

INSPECTION *7/8/25*

ACCEPTANCE *7/8/25*

Date Inspected:

Date Recieved:

☐ Inspected, verified and found in order as to quantity and specifications

☒ Complete

☐ Others: _____

☐ Partial

[Signature]
NOREVE JEAN M. AGAD
Inspector

[Signature]
VIVIAN V. BALBARINO
Head, Supply & Property Management