

Republic of the Philippines
VISAYAS STATE UNIVERSITY
Visca, Baybay City, Leyte
SUPPLY & PROPERTY MANAGEMENT

INSPECTION & ACCEPTANCE REPORT

Supplier: XPRT VENTURES INC.,	IAR No: 25-08-1020
P.O No.: BAC2-GOODS-25-14-005 (GF-CO)	Date: 2025-08-15
Office/Dept: DBS	345,000.00 GF-2024-10-02015

No.	UNIT	DESCRIPTION	Quantity		Date Received	Balance	Remarks
			Per P.O	Actual			
1	unit	Autoclave 100L Sterilizer, HYSC - AC 100L	1	1	2025-08-14 #: DR-3501/SI-0013	0	Completed

Specification:

Features:

Fuzzy function to prevent overheating

Accurate temperature controls via digital PID controller

Set time/temperature function

Error message display for temperature sensor disconnection

Alarm rings upon timer completion

Adjustable settings for discrepancies between real and displayed temperatures

Save and recall function to restore previous controller settings

Automatic process activation with a start button after setting temperature and time

Automatic air discharge and pressure elimination after reaching sterilizing temperature

Buzzer sounds at the end of the sterilizing period

Stable standing during cover opening/closing, aided by parking casters

Steam exit and valve located in a buried area for safety

Internal chamber made of STS #304 for durability

Auto and manual safety valves for safety

User-adjustable temperature limits

Temperature and Capacity:

Temperature Range: Max 125°C

Temperature Controller: Digital PID Controller

Temperature Sensor: Pt100

Pressure: 1.5 kg/cm²

Dimensions and Material:

External Size (WxDxH): approx. 600 x 710 x 1130 mm

External Material: Steel with powder coating

Basket: Stainless steel wire mesh (2 units)

Electric Requirements:

Voltage: 220 VAC, 50/60 Hz

Wattage: 4400 W

Safety Features:

Circuit breaker for electrical safety

Buried valve with a compact-designed power cable, safety pin, auto valve, and manual valve for maximum utilization

Manual safety valve: Easy to operate, indirect link, safe from hot steam

Silicone packing: Ultra-performance silicone packing

Wired basket: Wired basket with excellent thermal conductivity

Parking casters: Easy mobility and stable standing

Additional Features:

Threefold tempered safety glass windows for excellent insulation and easy observation

Buffering device to enhance door adherence for insulation

Upper vent for emitting moisture and gas after evaporation

Compact inside container for efficient use of space

Warranty: 3 yrs w/ 1 yr PMS

INSPECTION

Date Inspected: 8/15/15

☐ Inspected, verified and found in order
as to quantity and specifications

☐ Others: _____

ACCEPTANCE

Date Recieved: 8/15/15

☒ Complete

☐ Partial

NOREVE JEAN M. AGAD
Inspector

DOREEN B. ALBA
Head, Supply & Property Management