

Republic of the Philippines  
**VISAYAS STATE UNIVERSITY**  
Visca, Baybay City, Leyte  
**SUPPLY & PROPERTY MANAGEMENT**

## INSPECTION & ACCEPTANCE REPORT

Supplier: **NOVEAULAB ASIA CORP.**  
P.O No.: **BAC2-GOODS-25-20-093 (STF)**  
Office/Dept: **DSS**

IAR No: 25-09-1197  
Date: 2025-09-18  
**71,407.50**      **STF-2024-10-02113**

No.	UNIT	DESCRIPTION	Quantity		Date Received	Balance	Remarks
			Per P.O	Actual			
1	set	pH METER (benchtop)	1	1	2025-09-17 #: DR-009961/SI-7394	0	Completed

**Specification:**

- **Meter Type:** Benchtop
- **Display Type:** LCD
- **Connections:** pH (BNC), Temperature (2.5mm)
- **pH Range:** -2 to 16
- **pH Resolution:** 0.01-0.1
- **pH Accuracy:**  $\pm 0.01$  pH
- **Buffer Recognition:** NIST and USA
- **Calibration:** Up to 5 points
- **mV Range:** -2000 to 2000
- **mV Resolution:** 0.1 or 1 mV
- **mV Accuracy:**  $\pm 0.2$  mV or 0.05%, whichever is greater
- **Temperature Range:** 0 to 100°C
- **Temperature Compensation:** Automatic or Manual
- **Power Supply:** 100 to 240 VAC, 50 or 60 Hz
- **Dimensions (W x H):** 17.4625 cm x 6.985 cm (6.875 in x 2.75 in)
- **Description:** Benchtop meter with "All In One" pH capabilities
- **Warranty:** 3 years

INSPECTION

ACCEPTANCE

Date Inspected:

9/16/25

Date Recieved:

9/16/25

☒ Inspected, verified and found in order  
as to quantity and specifications

☐ Others: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

☒ Complete

☐ Partial

**NOREVE JEAN M. AGAD**

Inspector

**DOREEN B. ALBA**

Head, Supply & Property Management